



## QUALITY CERTIFICATE

### Selenotest ELISA Kit Version 4, C€

Colorimetric enzyme immunoassay for the quantitative measurement of human selenoprotein P in serum

Certificate valid for: LOT STE.16001

Product code: STE

Recommended storage: +2°C ... +8°C

Expiry date: 6 month after delivery

#### TEST KIT COMPOSITION:

No.	Identifier	Description	Product code	Lot No.	Packed
1	<span style="border: 1px solid black; padding: 2px;">SORB AB</span>	Assay plate	STE1	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE1.16001	<input type="checkbox"/>
2	<span style="border: 1px solid black; padding: 2px;">CAL 1 2 3 4 5 6 7 8</span>	Calibrator 1, 2, 3, 4, 5, 6, 7, 8	STE2	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE2.16001	<input type="checkbox"/>
3	<span style="border: 1px solid black; padding: 2px;">CONTROL L M H</span>	Control L, M, H	STE3	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE3.16001	<input type="checkbox"/>
4	<span style="border: 1px solid black; padding: 2px;">CONTROL B</span>	Control B	STE4	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE4.16001	<input type="checkbox"/>
5	<span style="border: 1px solid black; padding: 2px;">CONJ AB</span>	Detection antibody	STE5	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE5.16001	<input type="checkbox"/>
6	<span style="border: 1px solid black; padding: 2px;">CONJ EN</span>	Enzyme conjugate	STE6	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE6.16001	<input type="checkbox"/>
7	<span style="border: 1px solid black; padding: 2px;">BUF RCNS</span>	Reconstitution buffer	STE7	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE7.16001	<input type="checkbox"/>
8	<span style="border: 1px solid black; padding: 2px;">DIL AB</span>	Detection antibody diluent	STE8	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE8.16001	<input type="checkbox"/>
9	<span style="border: 1px solid black; padding: 2px;">DIL EN</span>	Enzyme conjugate diluent	STE9	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE9.16001	<input type="checkbox"/>
10	<span style="border: 1px solid black; padding: 2px;">DIL SPE</span>	Sample dilution buffer	STE10	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE10.16001	<input type="checkbox"/>
11	<span style="border: 1px solid black; padding: 2px;">BUF WASH 10x</span>	10× Washing buffer	STE11	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE11.16001	<input type="checkbox"/>
12	<span style="border: 1px solid black; padding: 2px;">SUBS TMB</span>	TMB substrate	STE12	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE12.16001	<input type="checkbox"/>
13	<span style="border: 1px solid black; padding: 2px;">SOLN STOP</span>	Stop solution	STE13	<span style="border: 1px solid black; padding: 2px;">LOT</span> STE13.16001	<input type="checkbox"/>
14	./.	Plate cover foil	STE14	<span style="border: 1px solid black; padding: 2px;">LOT</span> 16/151	<input type="checkbox"/>
15	./.	Instructions for use	./.	IFUv4_20170131_EN	<input type="checkbox"/>
16	./.	Product specification sheet	./.	PSS_STE.16001_EN	<input type="checkbox"/>
17	./.	This certificate	./.	QCC_STE.16001_EN	<input type="checkbox"/>

TEST KIT PACKED BY: \_\_\_\_\_

DATE: \_\_\_\_\_

QC testing: The quality testing was performed according to recommendations from the referred guidelines. Detailed results on page 2.

APPROVAL: The quality testing results meet the defined product specifications.

DATE: 2016-12-22



InVivo BioTech Services GmbH  
Neuendorfstr. 24a  
16761 Hennigsdorf  
Germany

phone: +49 3302 883 770  
fax: +49 3302 883 771  
email: [info@invivo.de](mailto:info@invivo.de)  
web: <http://www.invivo.de>

## QUALITY TESTING RESULTS:

## STANDARDS AND CONTROLS

CALIBRATOR/ CONTROL	Valid $E_{450\text{ nm}}$ range	Testing result $E_{450\text{ nm}}$	Valid concentration range SELENOP [ng/mL]	Testing result SELENOP [ng/mL]	Conclusion
CONTROL L	0.1066 – 0.1980	0.1523 ± 0.0122	16.1 – 21.7	19.0 ± 1.1	comply
CONTROL M	0.5486 – 1.0188	0.7837 ± 0.0651	63.2 – 85.6	75.1 ± 2.2	comply
CONTROL H	1.3553 – 2.5170	1.9362 ± 0.0792	233.0 – 315.2	271.8 ± 12.3	comply
CONTROL B	0.0306 – 0.0568	0.0437 ± 0.0008	n/a	N/A	comply
CAL 1	0.0416 – 0.0773	0.0594 ± 0.0035	5.0 – 9.2	7.3 ± 1.0	comply
CAL 2	0.0675 – 0.1254	0.0965 ± 0.0069	10.4 – 14.1	12.7 ± 0.6	comply
CAL 3	0.1653 – 0.3069	0.2361 ± 0.0196	23.1 – 31.2	27.2 ± 1.0	comply
CAL 4	0.3915 – 0.7270	0.5592 ± 0.0466	46.7 – 63.2	55.1 ± 0.9	comply
CAL 5	0.7682 – 1.4267	1.0975 ± 0.0781	91.0 – 123.2	107.8 ± 1.3	comply
CAL 6	1.1954 – 2.2201	1.7078 ± 0.0879	174.9 – 236.6	206.1 ± 3.6	comply
CAL 7	1.5237 – 2.8297	2.1767 ± 0.0964	341.2 – 461.6	390.6 ± 16.7	comply
CAL 8	1.6770 – 3.1144	2.3957 ± 0.1025	435.2 – 808.2	645.7 ± 57.9	comply

## ANALYTICAL PERFORMANCE PARAMETERS

PERFORMANCE PARAMETER	Specification	Testing result	Conclusion
Serum samples, accuracy, 4-PL	RE ± 25 %	+2.5 ± 11.3 %	comply
Serum samples, precision, intra-assay, 4-PL	RSD ≤ 20 %	8.3 ± 4.6 %	comply
Serum samples, precision, inter-assay, 4-PL	RSD ≤ 25 %	15.9 ± 1.8 %	comply
Serum samples, precision, inter-batch, 4-PL	RSD ≤ 30 %	20.4 ± 8.3 %	comply
Lower limit of detection, LLOD	< 1 mg/L	0.2 mg/L	comply
Upper limit of detection, ULOD	> 10 mg/L	20.5 mg/L	comply
Lower limit of quantitation, LLOQ	< 1 mg/L	0.4 mg/L	comply
Upper limit of quantitation, ULOQ	> 10 mg/L	13.2 mg/L	comply
Calibration curve, regression model, 4-PL	$r > 0.970$	0.993 ± 0.003	comply
Calibration curve, regression model, 5-PL	$r > 0.970$	0.995 ± 0.002	comply
Calibration curve, Calibrators and Controls, accuracy, 4-PL	RE ± 25 %	+1.2 ± 5.2 %	comply
Calibration curve, Calibrators and Controls, precision, 4-PL	RSD ≤ 25 %	5.9 ± 5.3 %	comply

REFERENCES: [1] US DEPARTMENT OF HEALTH AND HUMAN SERVICES, FDA, CDER and CVM. Guidance for the Industry: Bioanalytical Method Validation. Washington, DC, 2001; [2] INTERNATIONAL CONFERENCE OF HARMONIZATION (ICH) OF TECHNICAL REQUIREMENTS FOR THE REGISTRATION OF PHARMACEUTICALS FOR HUMAN USE. ICH Harmonised Tripartite Guideline. Validation of Analytical Procedures: Text and Methodology. Q2(R1). 2005; [3] DESILVA B., W. SMITH, R. WEINER, M. KELLEY, J. SMOLEC, B. LEE, M. KHAN, R. TACEY, H. HILL and A. CELNIKER. Recommendations for the bioanalytical method validation of ligand-binding assays to support pharmacokinetic assessments of macromolecules. *Pharm Res.* 2003, **20** (11), 1885-1900

**IMPORTANT NOTES:** Do not mix or interchange reagents from different Selenotest ELISA kit batches and do not use other reagents than provided by the kit or recommended by the instructions for use. Reagents from different Selenotest ELISA kits of the same batch can be mixed within the specified shelf life. Reagents from the same Selenotest ELISA kit batch were harmonised in order to ensure high reproducibility and optimal assay performance.